

# Technical Specifications

## Statement of Compliance

Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provisions of ITB Clause 3.1(a)(ii) and/or GCC Clause 2.1(a)(ii).

Please refer to the attached samples: Annex "A" for Face Masks and Annex "B" for Face Shields.

Description	Statement of Compliance
<b><u>Item 1- Face Masks</u></b> <ol style="list-style-type: none"> <li>1. Brand new.</li> <li>2. Non-woven, 3-ply Polypropylene fabric (ISO 1833; AATC 20A).</li> <li>3. Inner &amp; Middle layer should be water absorbent <math>\leq 5s</math> AATC79; Outer layer should water-repellent, at least 70 (ISO 4920; AATC 22).</li> <li>4. Flexible ear loop straps.</li> <li>5. Ergonomically and comfortable fit for all head sizes.</li> <li>6. Black color.</li> <li>7. With heat pressed 3 cms x 3 cms DAR logo located at the</li> </ol>	

<p>right side on lower cheek portion.</p> <p>8. Air Permeability should be 20-70cm<sup>3</sup>/s/cm<sup>2</sup> (ASTM D737; ISO 9237).</p> <p>9. Curved, molds around the bridge of the nose.</p> <p>10. A sample of the offer should be submitted during bid submission.</p> <p><b><u>Item 2- Face Shields</u></b></p> <p>1. Brand new.</p> <p>2. Hospital grade.</p> <p>3. Double size anti-fog, latex free, anti-static, distortion-free, no glare lens.</p> <p>4. Flexible frame with enough room for eyeglasses.</p> <p>5. Lightweight and comfortable to wear, quick and easy to don.</p> <p>6. Should wrap around the sides of the face and must extend below the chin.</p> <p>7. Ergonomically and comfortable.</p> <p>8. A sample of the offer should be submitted during bid submission.</p>	
--	--

**I hereby certify that the statement of compliance to the foregoing technical specifications and additional requirements are true and correct, otherwise, if found to be false either during bid evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.**

\_\_\_\_\_  
Name of Company/Bidder

\_\_\_\_\_  
Signature over Printed Name of  
Authorized Representative

\_\_\_\_\_  
Date

